

Sr. No.	Specification Compliance / Evaluation Parameters						
6	CENTRIFUGE	Bidder	THREE STAR	CAREVELL	SU Enterprises	SCIENTIFIC TECHNICAL	HOSPITECH HEALTHCARE
		Brand	PRESVAC	HTI DIAGNOSTIC	HETTICH	CENTURION SCIENTIFIC	DOMEL
		Model	DCS-16RTV	C8724	ROTOFIX32A	K-2015	CENTRIC 350
Country of Manufacturer		ARGENTINA	USA	GERMANY	UK	SLOVENIA	
Country of Origin of Product / Model		ARGENTINA	USA	GERMANY	UK	SLOVENIA	
Compliance with Defined quality Standards		Yes	Yes	Yes	Yes	Yes	
Specification Compliance feature wise		Remarks	Remarks	Remarks	Remarks	Remarks	
Specifications:		COUNTRY OF ORIGIN NOT USA/EUROPE/JAPAN,NO PAST PERFORMANCE AVAILABLE,NO DIRECT AUTHORISATION LETTER	Technically Acceptable	REMEMBER LAST SETTING FOR QUICK REPEAT RUNS NOT CONFIRMED FROM BROCHURE, ADJUSTABLE ACCELERATION AND DECELERATION NOT CONFIRMED,SUCTION CUP RUBBER FEET TO ELIMINATE VIBRATION NOT AVAILABLE	REQUIRED SPEED IS FROM 300 WHEREAS OFFERED STARTS FROM 500, SUCTION CUP RUBBER FEET TO ELIMINATE VIBRATION NOT AVAILABLE,NO PAST PERFORMANCE AVAILABLE	REMEMBER LAST SETTING FOR QUICK REPEAT RUNS NOT CONFIRMED FROM BROCHURE, SUCTION CUP RUBBER FEET TO ELIMINATE VIBRATION NOT AVAILABLE,STABILIZER IS CHINA MADE WHEREAS REQUIRED IS USA/EUROPE/JAPAN	
Technical Eligibility of Products:		NOT Eligible	Eligible	NOT Eligible	NOT Eligible	NOT Eligible	
Technical Eligibility of Firm:		Eligible	Eligible	Eligible	Eligible	Eligible	
Bid Status:		NON Responsive	Responsive	NON Responsive	NON Responsive	NON Responsive	

Zia ul Haq Tahir
Bio-Medical Engineer
Technical Officer, Health (Member)

Dr. Rafi Ahmed
Pathologist
PIMH, Lahore (Member)

Dr. Ashfaq Hameed Butt
DMS (General)
PIMH, Lahore (Member)

Dr. Tahir Pervaiz
Consultant Psychiatrist
PIMH, Lahore (Member)

Dr. Shahid Raza
Principal Medical Officer
PIMH, Lahore (Chairman)