

**PUNJAB INSTITUTE OF MENTAL HEALTH. TECHNICAL EVALUATION OF TENDER FOR THE PURCHASE OF LABORATORY EQUIPMENT UNDER  
ADP SCHEME 464 & 730 FOR THE FINANCIAL YEAR 2016-2017**

**EQUIPMENT NAME: Electrolyte Analyzer**

Sr. No	Evaluation Parameters	Scientific Technical Corporation	Gulf Marketing Intenational	Carevell	Sigma International
1	Complete Package / Tender	Yes	Yes	Yes	Yes
2	Original Receipt of Tender	Yes	Yes	Yes	Yes
3	Affidavit from Bidder	Yes	Yes	Yes	Yes
4	Bid Security	Yes	Yes	Yes	Yes
5	Bid Validity	Yes	Yes	Yes	Yes
6	Delivery Period	Yes	Yes	Yes	Yes
	Remarks:	Eligible	Eligible	Eligible	Eligible

Sr. No.	Evaluation Parameters	Scientific Technical Corporation	Gulf Marketing Intenational	Carevell	Sigma International
1	Exclusive Authorization / Sole Agent certificate by the manufacturer.	Yes but expired	Yes	Yes	Yes
2	Technical & Engineering capacity (As defined for the specific tender in specifications)	Yes	Yes	Yes	Yes
3	Certificate from the manufacturer about the after sales services through agent or itself (in case specifally demand n the specification)	Yes	Yes	Yes	Yes
4	Vendor past performance (in case of unsatisfactory performance details must be mentioned)	Yes	Yes	Yes	NOT RELEVANT TO QUOTED PRODUCT
5	Availability of relavent tools and testing / calibration equipment.	Yes	Yes	Yes	Yes
6	compliance of warranty as per tender	Yes	Yes	Yes	Yes
Remarks:		NOT Eligible	Eligible	Eligible	NOT Eligible

Sr. No.	Specification Compliance / Evaluation Parameters				
3	Electrolyte Analyzer	Bidder	Scientific Technical Corporation	Gulf Marketing International	Carevell
		Brand	JOKOH, CO. LTD.	Electra Medical Corporation	HTI DIAGNOSTIC
		Model	EX-Ds	Elite Plus	E-LYTE PLUS
Country of Manufacturer		japan	USA	USA	
Country of Origin of Product / Model Number		japan	TAIWAN	USA	
Compliance with Defined quality Standards		Yes	NO	Yes	
Specification Compliance feature wise		Remarks	Remarks	Remarks	
Specifications:		MANUAL TWO POINT CALIBRATION NOT AVAILABLE,2 OR MORE LEVELS OF QUALITY CONTROL NOT CONFIRMED FROM LEAFLET, SIMULTANEOUS DIGITAL DISPLAY OF RESULTS NOT VERIFIED FROM BROCHURE, ANALYSIS TIME SERUM/PLASMA/BLOOD/URINE NOT AVAILABLE,AUTHORITY LETTER EXPIRED.	Simultaneous digital display of result not confirmed from leaflet.automatic and manual two point calibration not available.	Technically Acceptable	
Technical Eligibility of Products:		NOT Eligible	Not Eligible	Eligible	
Technical Eligibility of Firm:		Eligible	Eligible	Eligible	
Bid Status:		NON Responsive	NON Responsive	Responsive	

**Zia ul Haq Tahir**  
Bio-Medical Engineer  
Technical Officer, Health (Member)

**Dr. Rafi Ahmed**  
Pathologist  
PIMH, Lahore (Member)

**Dr. Ashfaq Hameed Butt**  
DMS (General)  
PIMH, Lahore (Member)

**Dr. Tahir Pervaiz**  
Consultant Psychiatrist  
PIMH, Lahore (Member)

**Dr. Shahid Raza**  
Principal Medical Officer  
PIMH, Lahore (Chairman)