

PART- II

KNOCK DOWN CRITERIA - (VENDOR EVALUATION)

(To be evaluated by Technical Evaluation Committee)

(All evaluation parameters defined below are mandatory for compliance.)

Sr. No.	Evaluation Parameters	STARTEK INTERNATIONAL	MEDIQUIPS	BIOTECH SERVICES	SIGMA	CRYSTAL HEALTHCARE	GMED PVT LTD
1	Exclusive Authorization / Sole Agent Certificate by the Manufacturer	NO SUSPICIOUS AUTHORITY LETTER	YES	YES	YES	NOT FOR THE REGION	YES
2	Technical & Engineering capability(As defined for the specific tender in specifications)	YES	YES	YES	YES	YES	YES
3	Certificate from the Manufacturer about the after sales services through agent or itself (In case specifically demanded in the specifications)	YES	YES	YES	YES	NO	YES
4	Vendor Past performance (In case of unsatisfactory performance, details must be mentioned)	NO	YES	YES	YES	NO	YES
5	Availability of relevant Tools and Testing / Calibration Equipment	YES	YES	YES	YES	YES	YES
6	Compliance of Warranty as per tender	YES	YES	YES	YES	YES	YES
	Remarks:	(Not Eligibe for further evaluation of part-iii)	(Eligibe for further evaluation of part-iii)	(Eligibe for further evaluation of part-iii)	(Eligibe for further evaluation of part-iii)	(NOT Eligibe for further evaluation of part-iii)	(Eligibe for further evaluation of part-iii)

PART – III

KNOCK DOWN CRITERIA - PRODUCT EVALUATION

(All evaluation parameters defined below are mandatory for compliance.)

Item Sr.No	SPECIFICATION COMPLIANCE /EVALUATION PARAMETERS							
1	COMPANY	STARTEK INTERNATIONAL	MEDIQUIPS	BIOTECH SERVICES	SIGMA	CRYSTAL HEALTHCARE	GMED PVT LTD	
	DEFIBRILATOR	Brand	PROGETTI	NIHON KOHDEN	INNOMED	BEXEN CARDIO	US DEFIB	SCHILLER
		Model	RESCUE LIFE	TEC-5621K	CARDIO AID 360B	REANIBEX800	CARDIOSTART	DEFICART 4000
Country of Manufacturer		ITALY	JAPAN	HUNGARY	SPAIN	USA	FRANCE	
Country of Origin of Product/Model Number		_do_	_do_	_do_	_do_	_do_	_do_	
Compliance with defined quality standards		YES	YES	YES	YES	NO	YES	
Specification Compliance features wise:								
Specifications:		PEDIATRIC EXTERNAL PADDLES NOT OFFERED,PAPER SIZE 58MM INSTEAD OF 50MM, AUTHORITY LETTER AND PAST PERFORMANCE NOT AVAILABLE	AS PER SPECIFICATION	AS PER SPECIFICATION	CONTROL OF ENERGY CHARGING IS NOT AVAILABLE ON APEX PADDLE, BUILT IN BATTERY FOR ATLEAST 50 SHOCKS NOT CONFIRMED FROM LEAFLET	BATTERY BACK UP TIME NOT MENTIONED, SYNCRONISED OUTPUT WITH ECG NOT MENTIONED,ENERGY CONTROL ON APEX PADDLE NOT MENTIONED, QUALITY CERTIFICATE NOT AVAILABLE	ENERGY CHARGING ON FRONT PANEL IS NOT AVAILABLE	
Technical Eligibility of Product:		NO	YES	YES	NO	NO	NO	
Technical Eligibility of Firm:		YES	YES	YES	YES	YES	YES	
BID STATUS:		NON RESPONSIVE	Responsive	Responsive	NON Responsive	NON Responsive	NON Responsive	

Zia ul Haq Tahir

Bio-Medical Engineer

Technical Officer, Health (Member)

Dr. Tahir Pervaiz

Consultant Psychiatrist

Dr. Ashfaq Hameed Butt

DMS (General)

PIMH, Lahore (Member)

Dr. Shahid Raza

Principal Medical Officer

PIMH, Lahore (Member)

PIMH, Lahore (Chairman)