

**PUNJAB INSTITUTE OF MENTAL HEALTH. TECHNICAL EVALUATION OF TENDER FOR THE PURCHASE OF DENTAL EQUIPMENT UNDER
ADP SCHEME 464 & 730 FOR THE FINANCIAL YEAR 2016-2017**

BID EVALUATION SHEET

Name of the Equipment and Qty: DENTAL AUTOCLAVE

PART- I

KNOCK DOWN CRITERIA - (COMMERCIAL EVALUATION)

(To be evaluated by Purchase Department)

(All evaluation parameters defined below are mandatory for compliance)

Sr. No.	Evaluation Parameters	CAREVELL	ADVANCE SYSTEM	FORCARE	TOTAL TECHNOLOGY
1	Complete Package/Tender	YES	YES	YES	YES
2	Original Receipt of Tender	YES	YES	YES	YES
3	Affidavit from Bidder	YES	YES	YES	YES
4	Bid Security	YES	YES	YES	YES
5	Bid Validity	YES	YES	YES	YES
6	Delivery Period	YES	YES	YES	YES
Remarks:		(Eligibe for further evaluation of part-ii)	(Eligibe for further evaluation of part-ii)	(Eligibe for further evaluation of part-ii)	(Eligibe for further evaluation of part-ii)

PART- II**KNOCK DOWN CRITERIA - (VENDOR EVALUATION)****(To be evaluated by Technical Evaluation Committee)****(All evaluation parameters defined below are mandatory for compliance.)**

Sr. No.	Evaluation Parameters	CAREVELL	ADVANCE SYSTEM	FORCARE	TOTAL TECHNOLOGY
1	Exclusive Authorization / Sole Agent Certificate by the Manufacturer	YES	YES	YES	YES
2	Technical & Engineering capability(As defined for the specific tender in specifications)	YES	YES	YES	YES
3	Certificate from the Manufacturer about the after sales services through agent or itself (In case specifically demanded in the specifications)	YES	YES	YES	YES
4	Vendor Past performance (In case of unsatisfactory performance, details must be mentioned)	YES	YES	YES	YES
5	Availability of relevant Tools and Testing / Calibration Equipment	YES	YES	YES	YES
6	Compliance of Warranty as per tender	YES	YES	YES	YES
	Remarks:	(Eligibe for further evaluation of part-iii)	(Eligibe for further evaluation of part-iii)	(Eligibe for further evaluation of part-iii)	(Eligibe for further evaluation of part-iii)

PART – III

KNOCK DOWN CRITERIA - PRODUCT EVALUATION

(All evaluation parameters defined below are mandatory for compliance.)

Item Sr.No	SPECIFICATION COMPLIANCE /EVALUATION PARAMETERS					
1		COMPANY	CAREVELL	ADVANCE SYSTEM	FORCARE	TOTAL TECHNOLOGY
	DENTAL AUTOCLAVE	Brand	TRIDENT	TUTTNAUEER	FONA	MELAG
		Model	SOLE 17L	ELARA 9I	hygenius	VACUKLAVE 23 B +
Country of Manufacturer			ITALY	NEITHERLAND	SLOVAKIA REPUBLIC	GERMANY
Country of Origin of Product/Model Number			_do_	_do_	Italy	_do_
Compliance with defined quality standards			YES	YES	YES	YES
Specification Compliance features wise:						
Specifications:			AS PER SPECIFICATION	Don't have 12 Programe/Cycle. Moreover 2 Customizable cycle are not available. Option to lengthen drying cycle not available	Don't have 12 Programe/Cycle. Moreover 1 Customizable cycle insteaf of 2. Option to lengthen drying cycle not available.	have 7 Programe/Cycle. Instead of 12. Moreover 1 Customizable cycle insteaf of 2. Option to lengthen drying cycle not available
Technical Eligibility of Product:			YES	NO	NO	YES
Technical Eligibility of Firm:			YES	YES	YES	YES
BID STATUS:			RESPONSIVE	Non Responsive	Non Responsive	Non Responsive

Zia ul Haq Tahir
 Bio-Medical Engineer
 Technical Officer, Health (Member)

Dr. Ashfaq Hameed Butt
 DMS (General)
 PIMH, Lahore (Member)

Dr. Tahir Pervaiz
 Consultant Psychiatrist
 PIMH, Lahore (Member)

Dr. Shahid Raza
 Principal Medical Officer
 PIMH, Lahore (Chairman)