

PUNJAB INSTITUTE OF MENTAL HEALTH ,LAHORE

Auto Analyzer / Drug Analyzer

| Sr. No. | Specification Compliance / Evaluation Parameters | | | | | |
|---|--|--|---|---|---------------------------|----------------|
| 1 | Auto Analyzer / Drug Analyzer | Bidder | Merck | S Ejazudin & Co | Rawal Biomedical Services | CAREVELL |
| | | Brand | Vital Scientific | Erba Diagnostics | Beckman Coulter | HTI DIAGNOSTIC |
| | | Model | Selectra Pro-XL | ERBA XL-200 | AU-480 with ISE | BIOCHEM FC360 |
| Country of Manufacturer | | Netherlands | Germany, India | USA | USA | |
| Country of Origin of Product / Model Number | | Netherlands | Germany, India | Japan | USA | |
| Compliance with Defined quality Standards | | Yes | Yes | Yes | Yes | |
| Specification Compliance feature wise | | Remarks | Remarks | Remarks | Remarks | |
| Specifications: | | <p>Technically Not Acceptable</p> <ul style="list-style-type: none"> - No Build in Bar code Reader (reagents) - No Crash Prevention on Sample and Reagent Probes. - Disposable cuvette rotor after 10,000 tests. - 5 KVA On Line Sine wave UPS with Battery backup for 30 minutes (Emerson, Liebert, Chloride, MGE) not offered in Technical Offer - Compatible Laser Printer and RO system not offered in Technical Offer. | <p>Technically Not Acceptable.</p> <p>39 Positions for the Samples and QC/Calibrators while required are 50 positions. Database storage for minimum 100,000 results cannot be confirmed from leaflet.</p> <p>Remote factory support capability, rs232 interface, automatic flagging of results outside user defined data base and STAT function for priority sample cannot be confirmed from leaflet.</p> | <p>FDA 510K certificate not available, manufacturer is Probe with level detection and crash prevention system not available, water consumption is upto 20l/h whereas required is 1.5l/hour. It does not have the Remote Factory Support Capability (Internet access) not available, on board cooling from 2-8 degrees not confirmed from leaflet. R.O Unit is supplied locally furthermore make and model not mentioned where as required is USA/EUROPE/JAPAN</p> | Technically Acceptable | |
| Technical Eligibility of Products: | | Not Eligible | Not Eligible | Eligible | Eligible | |
| Technical Eligibility of Firm: | | Eligible | Eligible | Eligible | Eligible | |
| Bid Status: | | Non Responsive | Non Responsive | Non Responsive | Responsive | |

Zia ul Haq Tahir
Bio-Medical Engineer
Technical Officer, Health (Member)

Dr. Rafi Ahmed
Pathologist
PIMH, Lahore (Member)

Dr. Ashfaq Hameed Butt
DMS (General)
PIMH, Lahore (Member)

Dr. Tahir Pervaiz
Consultant Psychiatrist
PIMH, Lahore (Member)

Dr. Shahid Raza
Principal Medical Officer
PIMH, Lahore (Chairman)