

**PUNJAB INSTITUTE OF MENTAL HEALTH. TECHNICAL EVALUATION OF TENDER FOR THE PURCHASE OF ELECTROMEDICAL EQUIPMENT
UNDER ADP SCHEME 464 & 730 FOR THE FINANCIAL YEAR 2016-2017**

BID EVALUATION SHEET

Name of the Equipment and Qty: WALL MOUNTED PULSE OXIMETER

PART- I

KNOCK DOWN CRITERIA - (COMMERCIAL EVALUATION)

(To be evaluated by Purchase Department)

(All evaluation parameters defined below are mandatory for compliance)

Sr. No.	Evaluation Parameters	TRITECH	STARTEK INTERNATIONAL	BIOTECH SERVICES	CRYTSAL HEALTHCARE	GMED PVT LTD	AL KAREEM
1	Complete Package/Tender	YES	YES	YES	YES	YES	YES
2	Original Receipt of Tender	YES	YES	YES	YES	YES	YES
3	Affidavit from Bidder	NOT IN FAVOR OF PIMH	NOT IN FAVOR OF PIMH	YES	YES	YES	YES
4	Bid Security	YES	YES	YES	YES	YES	YES
5	Bid Validity	YES	YES	YES	YES	YES	YES
6	Delivery Period	YES	YES	YES	YES	YES	YES
Remarks:		(NOT Eligibe for further evaluation of part-ii	(Not Eligibe for further evaluation of part-ii	(Eligibe for further evaluation of part-ii	(Eligibe for further evaluation of part-ii	(Eligibe for further evaluation of part-ii	(Eligibe for further evaluation of part-ii

PART- II

KNOCK DOWN CRITERIA - (VENDOR EVALUATION)

(To be evaluated by Technical Evaluation Committee)

(All evaluation parameters defined below are mandatory for compliance.)

Sr. No.	Evaluation Parameters	TRITECH	STARTEK INTERNATIONAL	BIOTECH SERVICES	CRYSTAL HEALTHCARE	GMED PVT LTD	ALKAREEM
1	Exclusive Authorization / Sole Agent Certificate by the Manufacturer	YES	NO SUSPICIOUS AUTHORITY LETTER	YES	AUTHORITY LETTER NOT FOR THE REGION	YES	YES
2	Technical & Engineering capability(As defined for the specific tender in specifications)	YES	YES	YES	YES	YES	YES
3	Certificate from the Manufacturer about the after sales services through agent or itself (In case	YES	YES	YES	NO	YES	NO
4	Vendor Past performance (In case of unsatisfactory performance, details must be mentioned)	NO	NO	YES	NO	YES	YES
5	Availability of relevant Tools and Testing / Calibration Equipment	YES	YES	YES	YES	YES	YES
6	Compliance of Warranty as per tender	Not Offered	YES	YES	YES	YES	YES
	Remarks:	(Not Eligibe for further evaluation of part-iii)	(Not Eligibe for further evaluation of part-iii)	(Eligibe for further evaluation of part-iii)	(NOT Eligibe for further evaluation of part-iii)	(Eligibe for further evaluation of part-iii)	(NOT Eligibe for further evaluation of part-iii)

PART – III

KNOCK DOWN CRITERIA - PRODUCT EVALUATION

(All evaluation parameters defined below are mandatory for compliance.)

Item Sr.No	SPECIFICATION COMPLIANCE /EVALUATION PARAMETERS							
1		COMPANY	TRITECH	Startek International	BIOTECH SERVICES	CRYSTAL HEALTHCARE	GMED PVT LTD	AL KAREEM
	WALL MOUNTED PULSE OXIMETER	Brand	SCHILLER	Progetti	MEDLAB	MEDITECH	SCHILLER	ANDROMEDA
		Model	OXM PLUS	PG VITA	VITROMAP	M301	ARGOS OXM PLUS	VT900S
Country of Manufacturer			SWITZERLAND	ITALY	GERMANY	UK	SWITZERLAND	USA
Country of Origin of Product/Model Number			_do_	_do_	_do_	_do_	_do_	_do_
Compliance with defined quality standards			NO	YES	YES	YES	YES	NO
Specification Compliance features wise:								
Specifications:			AFFIDAVIT NOT IN FAVOR OF PIMH,PAST PERFORMANCE NOT AVAILABLE,WARRANTY NOT OFFERED,QUALITY CERTIFICATES NOT ATTACHED	PULSE RATE RANGE NOT AS PER TENDER SPECIFICATION,AUTHORITY LETTER UNCLEAR AND PAST PERFORMANCE NOT AVAILABLE	PULSE RATE NOT AS PER TENDER SPECIFICATION	PULSE RATE NOT AS PER TENDER SPECIFICATION,AUTHORITY LETTER NOT FOR THE REGION,AFTER SALE CERTIFICATE NOT AVAILABLE,VENDOR PAST PERFORMANCE NOT AVAILABLE	AS PER SPECIFICATION	PULSE STRENGTH PERFUSION INDICATION NOT CONFIRMED FROM LEAFLET,AFTERSALE CERTIFICATE FROM MANUFACTURER NOT AVAILABLE
Technical Eligibility of Product:			NO	NO	NO	NO	YES	NO
Technical Eligibility of Firm:			YES	YES	YES	YES	YES	YES
BID STATUS:			NON RESPONSIVE	Non Responsive	Non Responsive	NON Responsive	Responsive	NON Responsive

Zia ul Haq Tahir
 Bio-Medical Engineer
 Technical Officer, Health (Member)

Dr. Ashfaq Hameed Butt
 DMS (General)
 PIMH, Lahore (Member)

Dr. Tahir Pervaiz
 Consultant Psychiatrist
 PIMH, Lahore (Member)

Dr. Shahid Raza
 Principal Medical Officer
 PIMH, Lahore (Chairman)