

Minutes of Meeting of Grievance Committee under ADP schemes No. 730 & 464

A meeting of Grievance committee was held on 3-12-16 & 5-12-16 at 10 A.M vide letter no. in conference room of this institute to check the result of technical committee under ADP schemes no. 730 & 464 under the chairmanship of undersigned. The following member participated in the meeting:-

Dr. Ashique Hussain SMO Member

Dr. Mohsin Shaukat SMO Member

The proceeding was strated with the recitation from the Holy Quran. Total 13 firms have submitted grievances against the decision of technical committee. 01 firm namely Hospitech Healthcare has submitted its grievance late dated 03-12-2016 which committee has rejected on the ground that the time frame which was 01-12-2016has been lapsed. The detail of all others were as follow:-

Sr No	Name of bidder	Tender	Objection raised by T.Committee	Remarks
1	Makkays	Telephone exch	INSTALLAION OF WIRING/EXTENSION LINES NOT OFFERED.CONDITIONAL OFFER.System shall support VOIP (SIP or H.323). The IP card/module shall be integrated part of the System not available,,CPU 1+1 Redundant not available,Redundant central Processor complete in all respect in HOT standby mode not available.	The committee agreed with the stance of the company and cleared the objection. Hence technically accepted.
2	BIO Tech Services	Electro med ECG	Against the orient Defibrillator protection , Base line filter & LAN, WAN not available in model	The committee rejected the grievance of the company and uphold the decision of technical committee
3	Orient	Electro med ECG	Defibrillator protection , Base line filter & LAN, WAN not available in model	The committee agreed with the stance of the company and cleared the objection so technically accepted.
4	Gulf Marketing	Electro medic	Auto calibration not offered in technical bid	The committee rejected the grievance of the company and uphold the decision of technical committee because autocaliberation not offered in technical bid and leaflet/ broucher of the product is locally prepared.
5	Vital Care	Cardic monitor	Against carevell & Advance co. Past performance of the model	The committee rejected the grievance of the company and uphold the decision of

				technical committee
6	Al Raheem	Electro med equip	Cuff 3 size not offered	The committee accepted the stance of the company and accept its technical offer
7	Advance System	Dental unit	Against the carevell no sole authorization no past record	The committee rejected the grievance of the company and uphold the decision of technical committee
8	Apex Enterprises	Office furniture		The company was absent
9	Strongman	Laundry machinary	cycle time of washer not mentioned on leaflet,quoted model of tumbler is not microprocessor controlled,error code for troubleshot not available, no service program,no cool down time available,time cycle not confirmed from leaflet, flatwork ironer automatic linen detection and automatic cool down system not available,ironer speed adjustment starts from 1.5m/min instead of 1m/min,bearable temperature of belts 180°c instead of 200°c,	The committee rejected the the technical offer of pakage 01 of laundry while remaining items technically qualified
10	Qazafi Surgical	Hospital furnmiture	Over bed trolley top not of shesham(wooden)	The committee rejected the stance of the company and uphold the decision of the committee because the committee give its clarification about decision.
11	Sigma International	Lab Equip Elec analyz	Past performance not relevant	The committee has cleared the objection as the evidence was attached with bid
12	Noorani	Hospital furnmiture	Sample not as per specification small size of over bed trolley	The committee rejected the grievance of the company and uphold the decision of technical committee